



No. MMDSL/PROC-Eqpts/2023/(55)

Dated: 05/03/2024

CORRINGENDUM

Reference Tender Notice No: MMDSL/PROC-EQPTS/2023/(49) dated 11th.Feb.2024 for Supply & Installation of Various Equipment, the following corrigendum to the Tender notice are hereby notified through this to all concerned.

Sl. No	Page No./ Point No.	Amendment Required	Response/ Corrigendum
01	Page No.2 Eligibility Criteria Point No. 4 (f)-ii	Local Trade License Certificate clause should not be applicable as this is highly exclusive item (HDR Brachytherapy) involving Radioactive Source (Cobalt 60) with special Atomic Energy Regulatory Board approvals. This is an "one brand -onevendor" item. Kindly eliminate the clause for this item.	The bidder must submit an undertaking during tender submission, committing to furnish a valid Trade License Certificate from KHADC/JHADC/GHADC upon becoming the successful bidder. This certificate is required for non-tribal firms. Compliance with this requirement is mandatory for participation in the bidding process.
02	Page No.6 Performance Security Bond (PSB) Point No. 12 (a)	Kindly reduce Performance Security Deposit to 3% (three percent). Else high rate of 10% performance security deposit shall unnecessarily exaggerate the bidding price of the Bidders which shall not be economical for the Consignee / User.	The successful bidder must provide a Performance Security Deposit of 3% of the contract or purchase order value. This deposit is mandatory for successful bid acceptance.
03	Page No.6 Performance Security Bond (PSB) Point No. 12 (a)	Performance security in the form of PBG or FDR should be submitted in the warranty period + 3 months. For CMC performance security shall be submitted separately at the	Being a high cost value equipment, the CMC contract will commence right after the warranty period, hence clause remain unchanged.



Approved By Shri Ramkumar S (Managing Director) on 05/03/2024 07:00 AM
(The document is digitally approved and does not require any Seal or Signature in original)

		initiation of CAMC period.	
04	Page No. 23 Delivery: Point No. 23	Delivery period should be 270 days for Imported items. Note: Due to Force majeure conditions i.e. Post Covid World scenario & Russian Ukraine war, disturbance in middle east, the supply chain has been disrupted worldwide.	The delivery period is revised to 180 days. The bidder may submit valid reasons, as per tender conditions, for any delays encountered during the supply process.
05	Page No. 12 General Terms & Conditions Point No. 47	Since it is a high value item, Therefore must have provision of payment through 100% Indian LC (Indian Letter of Credit) - 90% on presentation of shipping documents & delivery at site & balance 10% on installation & commission.	The tender accepts the Letter of Credit (LC) however, LC payment is contingent on the availability of funds and compliance with LC terms. Non-fulfillment of the LC by the authority shall not account for supply delays.
06	Page No. 14 General Terms and Conditions Point No. 51(vi)-d	Preventive maintenance visit should be Half-yearly during CMC Period.	During the Comprehensive Maintenance Contract (CMC) period, preventive maintenance visits are scheduled half-yearly. However, upon Request during any particular time from the Medical Officer or Biomedical Engineer of MMDSL, preventive maintenance visits should be promptly arranged. Failure to respond promptly may result in penalties in accordance with the tender terms and conditions.
	Advanced ENT Operating Microscope	a) Magnification of 18X with Continuous Zoom. k) Integrated Video Camera Systems	NO CHANGE in specification) Continuous zoom is not mandatory. Companies quoting continuous zoom will



Approved By Shri Ramkumar S (Managing Director) on 05/03/2024 07:00 AM
(The document is digitally approved and does not require any Seal or Signature in original)

		<p>Integrated Beam Splitter also be accepted. 50:50 attached system CMOS medical grade HD Camera with HD Recorder with integrated Hard Disk o 1TB and HDMI input & Output Port. Record should be control easily from touchpad near to handgrip. 22" or more HD Monitor for Display.</p>	<p>k) Full integrated camera eliminates the need for any extra mount/cables that remains hanging around microscope head and restricts the maneuverability should be control easily from joystick near to handgrip and not touch pad Joystick can be controlled by the operating surgeon directly as and when required in the sterile condition. Touchpad is not sterile and requires an additional operator also.</p>
07	MICROSCOPE BODY	Total Magnification 48X"	NO CHANGE in specification Already mentioned 18x or better Companies quoting 48x will also be accepted.
08	<p>The following items are deleted from the tender</p> <ul style="list-style-type: none"> ● 1. Otoscope ● 2. Torch ● 3. GLUCOMETER ● 4. Glucometer Strip ● 5. Glucometer Lancet ● 6. Digital Haemoglobinometer ● 7. Slit Lamp ● 8. Trial Lens set (Paediatric) ● 9. Vision Testing drums ● 10. Illuminated Near Vision Test Drum ● 11. Suction Machine ● 12. Monitor ● 13. DIGITAL THERMOMETER ● 14. Radiant Warmer ● 15. ELECTRICAL SUCTION MACHINE ● 16. FOOT-OPERATED SUCTION MACHINE ● 17. Ultra sound cum Echocardiography machine for Padiatric & Neonate ● 18. ECG Machine (12 Lead) ● 19. OT instruments Set (Described Below) <ul style="list-style-type: none"> ● a) Artery Forcep St. & Curved S.S.6" ● b) Needle Holder 6 " 		



- c) Scissor Pointed Sharp 6"
- d) Mosquito Forcep 12.5 cm St. curved
- e) Epistomy Scissor 6"
- f) Stitch cutting Scissor 6"
- g) Allies Forcep 6"
- h) Mayo's Scissor 6"
- i) Sponge Holding Forcep 8"
- j) Cheatle Forcep 8"

- i) In light of the deleted items from the tender, Annexure-II, III/A, and III/B are hereby amended accordingly.
- ii) The bidders are to ensure that rate should be quoted as per the specified format, however the deleted items are to be excluded.
- iii) Last date and time for submission of Tender document: date is 12th/March/2024 at 11.30am
- iv) Tender opening date and time: 12th/March/2024 at 1:00pm
-

Please contact the office of the undersigned for any further inquiries or clarifications regarding the above corrigendum.

Other Terms and Conditions remain the same.

Shri Ramkumar S
Managing Director

Approved By Shri Ramkumar S (Managing Director) on 05/03/2024 07:00 AM
(The document is digitally approved and does not require any Seal or Signature in original)